



**CENTRE FOR MATERIALS FOR ELECTRONICS TECHNOLOGY (C-MET)**

(Scientific Society, Dept. of Information Technology, Govt. of India)

IDA PHASE – III, CHERLAPALLY, HCL (PO), HYDERABAD – 500 051

Phone: +91-40-27262437, 27267309, 27267006 Fax: +91-40-27261658

No. HD/PUR/SP-22/AC/29/2011-12/4486

January 19, 2012  
24

To

**Subject: Procurement of 'Air Compressor' – Changes in technical specifications – Regarding.**

Reference: Our Enquiry Letter No. HD/PUR/SP-22/AC/29/2011-12/4476 dated 18.01.2012

Sir,

In continuation to our above cited enquiry letter, you are requested to kindly note the following changes in the specifications for procurement of 'Oil Free Reciprocating Air Compressor (1 no.)'.

	<b>FOR</b>	<b>READ</b>
At Sl. No.:5 – Design Features	Working pressure at outlet of Integrated Refrigeration Type Air Dryer: <b>2</b> Kg./cm <sup>2</sup> .	Working pressure at outlet of Integrated Refrigeration Type Air Dryer: <b>8</b> Kg./cm <sup>2</sup> .
At Sl. No.:5 – Design Features – Control Panel	(c) Motor Overload Trip	(c) Motor Overload Trip/ <b>Short circuit/Earth Fault</b>

2. The other specifications and terms & conditions of Enquiry Letter cited above remain unaltered. You are requested to submit your quotation as early as possible by taking into consideration the above changes.

Thanking you,

Yours faithfully,

Sd/- xxx

(G.K.VENKATESAN)

Administrative Officer

[venkatesan@cmet.gov.in](mailto:venkatesan@cmet.gov.in)

[gkvenkatesh1966@gmail.com](mailto:gkvenkatesh1966@gmail.com)